

Action Plan

Service Name:	University Hospital Ayr
Service Provider:	NHS Ayrshire and Arran
Date of inspection:	28-29 April 2026

Recommendations	Action Planned	Timescale	Responsible Person	Progress	Date completed
a) UHA should review QA testing schedule for all diagnostic imaging equipment is in line or benchmarked against published guidance or where it deviates, this is part of a clear governance process.					
b) NHS A&A should ensure the results of all clinical audits undertaken at UHA are included in the reporting mechanism EP20 and or separately to the IRMER operational lead for inclusion in the IR(ME)R assurance mechanism.					
Requirements	Action Planned	Timescale	Responsible Person	Progress	Date completed
None.	-	-	-	-	-



Name: _____

Designation: _____

Signature: _____

Date: _____

In signing this form, you are confirming that you have the authority to complete it on behalf of the employer.